

CLARIFICATION#7

Request for Bids
Goods
(One-Envelope Bidding Process)

Country: Republic of Armenia

Project Name: Additional Financing for the Disease Prevention and Control
Loan No.: 9202-AM

RFB No: CR5/AF/RFB/B-G/001-22

Procurement of radiology diagnostic equipment, laboratory diagnostic equipment, surgery and hospital equipment, general hospital and polyclinic equipment and medical furniture for “Martuni Medical Center” CJSC of Gegharkunik Marz of the RA, including installation, start-up and training, comprising 5 lots

Dear Bidders,

In response to the request for clarification received on July 8, 2022, please find below (Section II - Bid Data Sheet, Clause ITB7.1) the respective response:

Q.: Please clarify when you specified Axial Rotation in the C-arm medical device requirements. not less than 400° for C-Arm, though 360° Axial Rotation is enough for full angles positioning

A.: The Axial Rotation for **Item#6 (Surgical X-Ray unit (C-Arm)) under Lot-1** shall be not less than 400° as specified in the Technical Specifications.

Q.: Please clarify, when you specified Rate of continues fluoroscopy in the requirements submitted to the C-Arm medical device not less than 30fps, though 30fps is high standard for cardiacs and brain interventional surgery, which should use suspended DSA C-arm. Mobile DSA C arm do not suggest to use these kind of surgeries.

A.: The rate of continues fluoroscopy for **Item#6 (Surgical X-Ray unit (C-Arm)) under Lot-1** shall be not less than 30fps as specified in the Technical Specifications.

Best regards,

Kristine Ghambaryan
Procurement Specialist

“Health Project Implementation Unit” State Agency
of the Ministry of Health of the RA

49/4 Komitas ave, 5th floor, 0051 Yerevan, RA (5th floor)

Tel. (+374 10) 29 75 36, (+374 10) 29 75 37

E-mail: info@healthpiu.am, procurement@healthpiu.am, kristine.ghambaryan@healthpiu.am

www.healthpiu.am